



Via Trionfale, 5952 - 00135 Roma - Tel. 06 355281 - Fax 06 35528319

Doc. Nr: 11080420  
Date issue : 08/04/2020  
Reference : Cov2 - Test  
Issued by: Laboratory Manager  
Approved by: Chief Operative Officer – Mr. Bruno Turchetta

### TO WHOM IT MAY CONCERN

The underwritten Casa di Cura VILLA STUART in Rome, is experimenting various advanced diagnostic systems for detection and contrast of SARS - CoV 2 diffusion in Italy. As part of this activity we tested the Dry Immunoassay System produced by Sichuan Xincheng Biological Co. Ltd, represented in Italy by CIMAF srl and MEDICA GROUP srl.

This system made by Sars-Cov-2 IgG/M antibody assay kit and AFS-1000 device is certified by CE and is under official certification by the Italian Health Public Institutions.

We tested on a cluster of 250 people potentially exposed to the infection for professional reasons. We have seen an ability to discriminate between patients with blower nose swab positive for SARS-CoV 2 RNA and control subjects higher than 95%.

It is a method that offers a good compromise between the lateral flow cards and the kits already existing on the market. It has the advantage of batch and sample traceability, objective fluorescence reading, ease of execution, small dimensions of the instrumentation which is easily transportable. This appears very useful in many contexts where the execution on large numbers on automated instruments is not indicated and certainly offers technical advantages over the card.

The above for the uses permitted by law.

Laboratory Manager

C.O.O.

UROSANITA' s.p.a.

Sede legale  
Piazza dei Caprettari, 70  
00186 Roma  
Tel. 06/66892680  
Soggetta a direzione  
e coordinamento della 3G s.p.a.

Sede amministrativa  
Via Casilina, 1049  
00189 Roma  
Tel. 06/231581  
Fax 06/2362288  
E-mail eurosanita@fiscal.it

Capitale sociale  
Euro 10.998.400,00 i.s.v.  
Reg. Imp. Roma 06726891002  
R.E.A. 987174  
C.F. e P.I. 06726891002